

TB-500 HPLC Report

Structure: TB-500

Lot NO :P170527-TB500

Column :250*4.6mm,Kromasil-C18-5um

Solvent A:0.1%TFA in 100%water


Solvent B:0.1%TFA in 100%acetonitrile

| Gradient : | A | B |
|------------|------|------|
| 0.1min | 83% | 17% |
| 25.0min | 58% | 42% |
| 25.1min | 0% | 100% |
| 30.0min | stop | |

Flow rate:1.0ml/min

Wavelength(nm):220

Volume :5ul



| Peak No | Ret Time | Height | Area | Conc. |
|---------|----------|------------|-------------|---------|
| 1 | 4.670 | 2208.462 | 16468.801 | 0.2202 |
| 2 | 6.295 | 1203.221 | 9089.100 | 0.1215 |
| 3 | 10.305 | 4262.055 | 9764.494 | 0.1305 |
| 4 | 10.545 | 775687.500 | 7398914.000 | 98.9122 |
| 5 | 13.533 | 2523.888 | 21007.000 | 0.2808 |
| 6 | 20.993 | 3534.210 | 25037.297 | 0.3347 |

100.00