

Thymalin HPLC Report

Structure: Thymalin

Lot NO :P170320-Thymalin

Column :250*4.6mm,Kromasil-C18-5um

Solvent A:0.1%TFA in 100%water


Solvent B:0.1%TFA in 100%acetonitrile

Gradient :	A	B
0.1min	98%	2%
25.0min	73%	27%
25.1min	0%	100%
30.0min	stop	

Flow rate:1.0ml/min

Wavelength(nm):220

Volume:10ul



Peak No.	Ret Time	Height	Area	Conc.
1	11.005	1214.525	8562.113	0.1110
2	11.842	3984.961	26049.939	0.3379
3	11.985	662837.000	7625306.500	98.8962
4	12.297	13062.425	50496.293	0.6549
				100.00